

MEDTRONIC 770G PUMP UPGRADE & CGM TRAINING

In order to be prepared for training, please take the time to do the following to ensure that together we can optimize your learning experience:

1. Please open your box(s), take everything out and look for your Medtronic MiniMed 770G Training guides. Please read these in detail, so that you are prepared.
2. Please go to <https://www.medtronic.com/ca-en/diabetes/home/support/product-support/minimed-770g-training.html> to access the online learning modules. Please complete the MiniMed 770G training modules prior to your training. Please place batteries in your pump, enter the time and date, your current insulin: carb ratios, correction factors and targets, and please practice button-pushing before your appointment.
3. If you will be starting CGM at the time of your pump upgrade training please make sure you have ordered and received your Guardian Transmitter, that you have charged it, and that you have ordered and received your sensors. If you require sensors, please order them by calling **1-800-284-4416**. Many leave this until the last minute and find that they do not have what they need for training.
4. Please go to <http://www.carelink.minimed.eu> and set up your Medtronic CareLink Personal account, to access our web-based data management tool, which will help you and your healthcare professional gain insights from your glucose trends.

In preparation for the training on your new pump, these are the settings you will need from the pump you are using now:

Insulin: carbohydrate ratios (ICR)

Insulin sensitivity or correction factor (ISF)

Basal programs

Basal rates

BG Targets

Active Insulin time (Insulin action)

Low reservoir

Max basal

Max bolus

Auto off (Auto suspend)

Please contact Medtronic Technical Support at 1-800-284-4416 if you have any questions as you complete the preparation for training.